

with an "A" type scale. It has been

requirements set out in P.M.G. Specific

OPERATING INSTRUCTIONS AND

INPUT:- The DESCRIPTION is greater

DESCRIPTION

Frequency range from 30 cycles to 15 kil

for

ent can satisfactorily bridge levels fro

VOLUME INDICATOR


terminal strip and the looks on the tr

V.U. METE meter calibrated fr

an accordance with Schedule C.7248, and

rolled by two attenuators so that the me

between -20 V.U. and +3 V.U.



Manufactured in Australia by

TRANSMISSION PRODUCTS PTY. LTD.

NORTH SYDNEY

OPERATING INSTRUCTIONS AND
DESCRIPTION

for

VOLUME INDICATOR

Schedule:	C.7188
Contract:	19396.
Item No:	5.
Serial No:	

HANDBOOK.

VOLUME INDICATOR.

P.M.G. Specification No.
722A

GENERAL DESCRIPTION:

This instrument is a two stage amplifier employing two 6 AU 6 Valves with the output terminated with a 7500 ohm V.U. Meter which conforms to the requirements set out in P.M.G. Schedule C.7248, with an "A" type scale. It has been designed to fully meet the requirements set out in P.M.G. Specification 722A.

INPUT:- The input impedance is greater than 25,000 ohms over a frequency range from 30 cycles to 15 kilocycles. The instrument can satisfactorily bridge levels from -40 V.U. to +33 V. Access to the input of the instrument is available at both the terminal strip and the jacks on the front panel.

V.U. METER:- The meter is calibrated from -20 V.U. to +3 V.U. in accordance with Schedule C.7248, and the input level is controlled by two attenuators so that the measured output level will lie between -20 V.U. and +3 V.U.


ATTENUATORS:- The first attenuator is a three step "T" type attenuator calibrated -10, 0, +20 V.U. The second attenuator is a ten step potentiometer calibrated from -10 to +10 V.U. in 2 V.U. steps, the combined accuracy of these two attenuate is better than ± 0.2 V.U. at any setting when measured at 1 Kc/s.

FREQUENCY RESPONSE:- The variation in gain for frequencies ranging from 30 cycles to 15 Kc/s, is less than ± 0.5 db from the gain measured at 1 Kc/s.

CALIBRATION:- A potentiometer marked "Calibrate", is included in the circuit so that aging of the tubes can be compensated for, and an exact adjustment can be made.

POWER SUPPLY:- This instrument is designed to operate from 132V. D.C. Anode supply, and a 24V D.C/A.C. Filament supply. A filament on/off switch is provided so that the instrument may be switched off when not in use. The off position of this switch provides a protective shunt across the V.U. Meter.

ITEM No.
REF.
OPERATIONS



INPUT IMPEDANCE
The input impedance shall be high (greater than 25000Ω at 1 kcs)

FREQUENCY RESPONSE
The frequency response shall not differ by more than 0.5 db from the gain at 1000 g/s over the frequency range extending from 30 g/s to 15 kcs.

RANGE
The operating range shall be 73 v.u. and shall range from -40 v.u. to +33 v.u.

ACCURACY
The maximum error at any setting of the attenuator shall not exceed ±0.2db when measured at 1000 g/s.

1	ORIGINAL
ISSUE	CHANGE
CHANGES	

MATERIAL		DRAWN		TRANSMISSION PRODUCTS PTY. LTD. COMMUNICATION ENGINEERS DENISON STREET, NORTH SYDNEY, AUSTRALIA
QUALITY		TRACED		
STOCK SIZE		CHECKED		
WEIGHT, 1,000		PASSED		
LENGTH, 1,000		APPROVED		
SHEETS, 1,000		DATE		SCHEMATIC VOLUME INDICATOR
ITEM FINISH				
THIRD ANGLE PROJECTION — DO NOT SCALE DRAWING — REPORT ERRORS				DRAWING NUMBER
SCALE: —				UNLESS SPECIFIED TOLERANCE ON FRACTIONS IS: — DECIMALS IS: —
				MO9-147-1